

Time Frame	Anticipated Service
<b>Year One</b>  First two cleaning and calibrations No Charge when analyzer sent back within first year. Shipping and Handling costs not included	No Charge
<b>Year One to Two</b>  Calibrate every 5-6 months or as required based on use Replacement of O2 Sensor	2 Hours labour Calibration Gases Sensor if required
<b>Year Two to Three</b>  Calibrate every 5-6 months or as required based on use	2 Hours labour Calibration gases
<b>Year Three to Four</b>  Calibrate every 5-6 months or as required based on use Possible replacement of Toxic gas sensors such as CO/NO/SO2	2 Hours labour Calibration gases

**Notes:**

Consumables such as batteries for printer, printer paper and analyzer filters are not considered.

Analyzer not subjected to misuse or handling such as excessive CO measurements or excessive cold temperatures.

All service and calibration work is provided by KWE Technologies Group. Contact for parts and service pricing.